

Urgent Field Safety Notice

Product name: Affinis Fracture, Inverse, and Short implants with Ti/CaP coating

FSCA ID no.: FSCA 16/02

Type of action: Replacement of products bearing the wrong expiry date

Bettlach, 31 May 2016

Issued by: Mathys Ltd Bettlach

Addressee: OR management

CC: Purchasing department

Affected products:

Product	Item number	Item description	Batch numbers
	60.21.0000	Affinis Fracture Central Part 1	see Annex
	60.21.0001	Affinis Fracture Central Part 2	see Annex
	60.30.3150	Affinis Inverse Metaglene	see Annex
	60.30.3151	Affinis Inverse Metaglene Revision	see Annex
	60.30.6390	Affinis Fracture Inverse 39+0	see Annex
	60.30.6393	Affinis Fracture Inverse 39+3	see Annex
	60.30.6420	Affinis Fracture Inverse 42+0	see Annex
	60.30.6423	Affinis Fracture Inverse 42+3	see Annex



	62.34.0010	Affinis Short Stem 1	see Annex
	62.34.0011	Affinis Short Stem 2	see Annex
	62.34.0012	Affinis Short Stem 3	see Annex
	62.34.0013	Affinis Short Stem 4	see Annex
	62.34.0014	Affinis Short Stem 5	see Annex
	62.34.0015	Affinis Short Stem 6	see Annex

Ladies and Gentlemen

Mathys Ltd Bettlach hereby notifies you of a voluntary Field Safety Corrective Action (FSCA) affecting the item and batch numbers of Affinis Fracture, Inverse, and Short implants listed in the Annex. Our records indicate that you have received one or several of the affected batches and are therefore subject to this safety-relevant action.

Description of the problem

Mathys Ltd Bettlach has detected an inconsistency in the shelf-life of implants coated with Titanium/Calcium Phosphate (Ti/CaP). As a preventive measure, the shelf-life of the Ti/CaP coated implants has been reduced from previously ten years to a new shelf-life of five years. For this reason, all the stocks of affected implants have been relabelled accordingly.

Potential hazards

The coating of Titanium/Calcium-Phosphate (Ti/CaP) coated implants that exceed the shelf-life of five years might undergo a chemico-physical change. There is an unlikely probability of less efficient osseointegration within the first six to twelve weeks after implantation.

The potential chemico-physical change does <u>not</u> have an influence on aseptic loosening of components with Ti/CaP coating twelve or more weeks after implantation.

The first batches which are labelled with a shelf-life of ten years will exceed the new, corrected shelf-life of five years in August 2016.



Immediate actions to be taken by the customer:

- Read this Field Safety Notice carefully and make sure that all relevant departments and positions are informed of its content.
- Immediately separate out all the products carrying the item and batch numbers indicated above.
- A Mathys representative will contact you in order to support you with this procedure and organise the replacement.
- Third parties to whom affected products have been forwarded must be informed and instructed accordingly.
- Please complete the enclosed confirmation form and return it to the address indicated or hand it
 over to your Mathys representative. (This will stop Mathys from sending you further reminders
 concerning this FSCA.)
- The present Field Safety Notice must be observed until the action has been completed within your organisation. Please keep a copy of this Field Safety Notice.
- For questions regarding the replacement, please contact the Mathys representative responsible for your organisation or your local Mathys office.
- For questions regarding this Field Safety Notice, please contact us at the following address: vigilance@mathysmedical.com


Information on materiovigilance:

The competent national authorities have been notified of this action.

Please notify Mathys Ltd Bettlach of any adverse event in connection with the affected product or any other Mathys product. You can report adverse events to Mathys at vigilance@mathysmedical.com or via your local Mathys office.

We very much regret the inconvenience caused to you. If you have any further queries, please do not hesitate to contact us.

Mathys Ltd Bettlach





Confirmation form FSCA 16/02

Urgent Field Safety No	<u>otice</u>			
Product name:	Affinis Fracture, Inverse, and Short implants with Ti/CaP coating			
FSCA ID no.:	FSCA 16/02			
Type of action:	Replacement of products bearing the wrong expiry date			
	Confirmation of receip	t		
Please complete:				
Customer no.				
Hospital				
Post code, town				
Contact (Name/position)				
By completing and retu Safety Notice:	rning the present form I confirm that I h	ave received and read this Field		
Our stocks do not con	tain any affected products			
Our stocks contain the	e following affected products:			
Item number	Batch	Number of units		
Place/date:	Signature:			
Please return this forme-mail: fax:	m by e-mail or fax to the following ac	ddress:		