

27 September 2017

To: Surgeons/ Hospitals

Subject: **URGENT MEDICAL DEVICE FIELD SAFETY NOTICE (REMOVAL)**

Reference: **FA 2017-04 (ZFA2017-332)**

Affected Product: Specific Hip and Trauma instruments

Material Number	Description	Material Group
01.00069.409	Drill cavity ø 9 mm	Instrument Hip
01.00069.410	Drill cavity ø 10 mm	Instrument Hip
01.00069.411	Drill cavity ø 11 mm	Instrument Hip
01.00069.412	Drill cavity ø 12 mm	Instrument Hip
01.00069.413	Drill cavity ø 13 mm	Instrument Hip
01.00069.414	Drill cavity ø 14 mm	Instrument Hip
01.00069.415	Drill cavity ø 15 mm	Instrument Hip
01.00069.416	Drill cavity ø 16 mm	Instrument Hip
01.00069.417	Drill cavity ø 17 mm	Instrument Hip
75.80.04	Flexible shaft	Instrument Hip
110.44.150	Flexible shaft for intramedullary reamer heads, max. depth 440 mm, ø 9–12.5 mm	Instrument Trauma Instrument Trauma
110.44.155	Flexible shaft for intramedullary reamer heads, max. depth 440 mm, ø 13–19 mm	Instrument Trauma
110.44.207	Flexible intramedullary reamer monobloc, front cutting, ø7mm	Instrument Trauma
110.44.208	Flexible intramedullary reamer monobloc, front cutting, ø8 mm	Instrument Trauma
110.44.209	Flexible intramedullary reamer monobloc, front cutting, ø9 mm	Instrument Trauma
02.00020.040	Drill ø 13 mm with flexible shaft	Instrument Trauma

Table 1: Affected products

Zimmer GmbH is conducting a medical device field action (removal) for specific hip and trauma instruments (described in table 1). These instruments are part of an outdated technology.

Consequently there is a potential that the instruments may not be adequately cleaned when utilizing the standard cleaning instructions. If an instrument is not adequately cleaned, this could result in infection and subsequent complications. As a result, the devices are being removed and as required being replaced with alternative instruments (already available) that can be adequately cleaned utilizing the standard cleaning instructions (see attachment 2 for the replacements).



Our records indicate you may have received one or more of the affected products. Please note that we have no records of reports related to infection related to this issue.

Surgeon/ Hospital Responsibilities:

1. Review this notification for awareness of the contents.
2. Assist your Zimmer Biomet sales representative to quarantine all affected product.
3. Your Zimmer Biomet sales representative will remove the affected product from your facility.
4. There are no specific patient monitoring instructions related to this field action that are recommended beyond your existing follow up schedule.
5. Complete Attachment 1 – Certificate of Acknowledgement.
 - a. Return a digital copy to fieldaction.emea@zimmerbiomet.com.
 - b. Retain a copy of the Certificate of Acknowledgement with your field action records in the event of a compliance audit of your documentation.
6. If after reviewing the notice you have further questions or concerns please contact your Zimmer Biomet representative.

Other Information

This voluntary medical device Field Safety Notice was reported to all relevant Competent Authorities and the related Notified Body as required under the applicable regulations for Medical Devices as per MEDDEV 2.12-1 in Europe.

Please keep Zimmer Biomet informed of any adverse events associated with this product or any other Zimmer Biomet product by emailing winterthur.per@zimmerbiomet.com or to your local Zimmer Biomet contact.

Please be aware that the names of user facilities notified are routinely provided to the Competent Authorities for audit purposes.

The undersigned confirms that this notice has been delivered to the appropriate Regulatory Agencies.

We would like to thank you for your co-operation in advance and regret any inconveniences caused by this field action.

Sincerely,

A handwritten signature in blue ink that reads 'Matthias Bürger'. The signature is written over a horizontal line.

Matthias Bürger

Zimmer Biomet Vice President QARC EMEA



ATTACHMENT 1
Certificate of Acknowledgement

FA2017-04 (ZFA2017-332)

By signing below, I acknowledge that the required actions have been taken in accordance with the Field Safety Notice.

Hospital Facility **Surgeon** (Please check one as applicable)

Printed Name: _____ **Signature:** _____

Title: _____ **Telephone:** () _____ - _____ **Date:** ____ / ____ / ____

Facility Name: _____

Facility Address: _____

City: _____ **ZIP:** _____ **Country:** _____

Note: This form must be returned to Zimmer Biomet before this action can be considered closed for your account. It is important that you complete this form and email a copy to: fieldaction.emea@zimmerbiomet.com.

Product Reference	Number of products returned



ATTACHMENT 2- List of replacement products

Material Number	Description	Material Group	Implant System	Replacement reference	Description
1.00069.409	Drill cavity ø 9 mm	Instr. Hip	Optan Stem System	00-2228-009-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 9.0MM DIA
1.00069.410	Drill cavity ø 10 mm	Instr. Hip	Optan Stem System	00-2228-010-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 10.0MM DIA
1.00069.411	Drill cavity ø 11 mm	Instr. Hip	Optan Stem System	00-2228-011-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 11.0MM DIA
1.00069.412	Drill cavity ø 12 mm	Instr. Hip	Optan Stem System	00-2228-012-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 12.0MM DIA
1.00069.413	Drill cavity ø 13 mm	Instr. Hip	Optan Stem System	00-2228-013-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 13.0MM DIA
1.00069.414	Drill cavity ø 14 mm	Instr. Hip	Optan Stem System	00-2228-014-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 14.0MM DIA
1.00069.415	Drill cavity ø 15 mm	Instr. Hip	Optan Stem System	00-2228-015-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 15.0MM DIA
1.00069.416	Drill cavity ø 16 mm	Instr. Hip	Optan Stem System	00-2228-016-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 16.0MM DIA
1.00069.417	Drill cavity ø 17 mm	Instr. Hip	Optan Stem System	00-2228-017-00	PRESSURE SENTINEL FLEXIBLE IM REAMER 17.0MM DIA
5.80.04	Flexible shaft	Instr. Hip	various Acetabular shell systems	00-8790-007-05	Modular Flexible Shaft (tri-shank)
10.44.150	Flexible shaft for intramedullary reamer heads, max. depth 440 mm, ø 9–12.5 mm	Instr. Trauma	Sulzer Medica reamer instruments for intramedullary nails for femur and tibia	no replacement	n/a
10.44.155	Flexible shaft for intramedullary reamer heads, max. depth 440 mm, ø 13–19 mm	Instr. Trauma	Sulzer Medica reamer instruments for intramedullary nails for femur and tibia	no replacement	n/a
10.44.207	Flexible intramedullary reamer monobloc, front cutting, ø7mm	Instr. Trauma	Sulzer Medica reamer instruments for intramedullary nails for femur and tibia	00-2228-007-00	Pressure Sentinel 7.0mm Flexible Reamer
10.44.208	Flexible intramedullary reamer monobloc, front cutting, ø8 mm	Instr. Trauma	Sulzer Medica reamer instruments for intramedullary nails for femur and tibia	00-2228-008-00	Pressure Sentinel 8.0mm Flexible Reamer
10.44.209	Flexible intramedullary reamer monobloc, front cutting, ø9 mm	Instr. Trauma	Sulzer Medica reamer instruments for intramedullary nails for femur and tibia	00-2228-009-00	Pressure Sentinel 9.0mm Flexible Reamer
2.00020.040	Drill ø 13 mm with flexible shaft	Instr. Trauma	Sirus Implant System	no replacement	n/a