



[FDA Home](#)<sup>3</sup> [Medical Devices](#)<sup>4</sup> [Databases](#)<sup>5</sup>

## Class 1 Device Recall Roche CoaguChek XS System Prothrombin time test



6 510(k) | De Novo<sup>8</sup> | Registration & Listing<sup>9</sup> | Adverse Events<sup>10</sup> | Recalls<sup>11</sup> | PMA<sup>12</sup> | HDE<sup>13</sup> | Classification<sup>14</sup> | Standards<sup>15</sup>  
 CFR Title 21<sup>16</sup> | Radiation-Emitting Products<sup>17</sup> | X-Ray Assembler<sup>18</sup> | Medsun Reports<sup>19</sup> | CLIA<sup>20</sup> | TPLC<sup>21</sup>

[New Search](#)

[Back to Search Results](#)

### Class 1 Device Recall Roche CoaguChek XS System Prothrombin time test



<b>Date Initiated by Firm</b>	September 13, 2018
<b>Date Posted</b>	November 02, 2018
<b>Recall Status<sup>1</sup></b>	Open <sup>3</sup> , Classified
<b>Recall Number</b>	Z-0360-2019
<b>Recall Event ID</b>	<u>81093</u> <sup>23</sup>
<b>510(K)Number</b>	<u>K093460</u> <sup>24</sup> <u>K071041</u> <sup>25</sup> <u>K180684</u> <sup>26</sup> <u>K170960</u> <sup>27</sup>
<b>Product Classification</b>	Test, time, prothrombin <sup>28</sup> - <b>Product Code GJS</b> <sup>29</sup>
<b>Product</b>	CoaguChek <sub>z</sub> XS System Prothrombin time test: CoaguChek XS PT Test 2X24 Strips, Catalog Number: 04625315160; CoaguChek XS PT Test 6 Strips, Catalog Number: 04625374160; CoaguChek XS PT Test 24 Tests, Catalog Number: 07797826160;
<b>Code Information</b>	Lot numbers: 28124111, 28124121, 28631911, 28631921, 28631924, 28632021, 28632213, 28632312, 28632412, 29415113, 29415123, 29494221, 29494312, 29494613, 29494711, 29778721, 29779012, 29779213, 29779214, 30497213, 30497311, 30497413, 30497423, 30497515, 31404314, 31404821, 32264116, 32264212, 32264316, 32264317, 32264411, 32264421, 33045913, 33046011, 33046113, 33046312, 33046314, 33046321, 33046322, 33449612, 33449712, 33449723, 33449817
<b>Recalling Firm/ Manufacturer</b>	Roche Diagnostics Corporation 9115 Hague Rd Indianapolis IN 46256-1025
<b>For Additional Information Contact</b>	SAME 317-5214343
<b>Manufacturer Reason for Recall</b>	Abnormally high INR test results with the affected CoaguChek test strips
<b>FDA Determined Cause<sup>2</sup></b>	Nonconforming Material/Component
<b>Action</b>	Roche issued URGENT MEDICAL DEVICE RECALL AMENDMENT letters dated 11/5/18 advising customers patients, HCPs, IDTFs, HCP distributors and HCP/patient distributors. Distributors are advised to take the following actions: - Read the Urgent Medical Device Recall Amendment for Healthcare Professionals and Clinics Furnishing PT/INR Test Strips to Patients for Home Testing Use (TP-00454). - Distribute a copy of the Urgent Medical Device Recall Amendment TP-00454 to healthcare professionals to whom you have directly shipped affected test strips. - Stop distribution of all affected CoaguChek XS PT Test Strip lots. - Complete all sections of the fax form (TP-00462) and fax it to number 1-888-627-2279 or email it to roche3866@stericycle.com. - Return all affected product following the instructions outlined in this Urgent Medical Device Recall Amendment. Independent Diagnostic Testing Facilities should take the following action: - Read the Urgent Medical Device Recall Amendment for CoaguChek Patients (TP-00453) and Urgent Medical Device Recall Amendment for Healthcare Professionals and Clinics Furnishing PT/INR Test Strips to Patients for Home Testing Use (TP-00454). - Distribute a copy of the Urgent Medical Device