



رقم المحفوظات: ٨١٢٥
رقم الصادر: ١٣/١/٢٥٢٥٧
بيروت، في: ٣٠ تموز ٢٠١٣

جانب نقيب الاطباء في الشمال/طرابلس

الموضوع: إشعار بمتابعة جهاز طبي مغروس

Hydrocephalic shunt, Codman Certas Programmable Valve

الجهاز المعنى بالمتابعة:

- Hydrocephalic shunt, Codman Certas Programmable Valve
- Trade Mark: DePuySynthes
- Local Representative:

بناء على التوصية الصادرة عن الشركة المصنعة والتي تشير الى عدم ملاءمة الصنف الوارد أعلاه لشروط التصوير بجهاز الرنين المغناطيسي، نرجو منكم متابعة هذا الموضوع مع الاطباء الاختصاصيين.

مرفق ربطاً:


- التوصية الصادرة عن الشركة المصنعة

يبلغ:


- دائرة البرامج والمشاريع
- المستشفيات الحكومية
- المحفوظات

مدير عام الصحة

د. وليد عمار



CODMAN NEURO



Urgent Field Safety Notice (FSN)

Product Name: CODMAN[®] CERTAS[™] Programmable Valves

FSCA-identifier: Our ref: COM-009386

Type of Action: Field Safety Notice and Field Safety Corrective Action

Date: May 15, 2013

Attention: Trust Chief Executives, Surgery,
Theatre and Neurosurgical General Managers – Private Sector Hospitals

Affected devices: *Codman Certas Programmable Valves*

Part numbers:

Code	Description
82-8800	In Line Valve Only
82-8801	In Line Valve with Catheter and Accessories
82-8802	In Line Valve with Unitized Catheter and Accessories
82-8803	In line Valve with Unitized BACTISEAL [®] Catheter and Accessories
82-8804	In Line Valve only with SIPHONGUARD [®] Device
82-8805	In Line Valve with SIPHONGUARD [®] Device, Catheter and Accessories
82-8806	In Line Valve with SIPHONGUARD [®] , Unitized Catheter and Accessories
82-8807	In Line Valve with SIPHONGUARD [®] Device, Unitized BACTISEAL [®] Catheter and Accessories

Lot No/Serial No: See attached List

Reason for this Recall:

This recall is being initiated because our testing has shown that, in a small percentage of valves, the following conditions may exist:

1. **The MRI resistance feature may not always operate properly.** If this occurs, it could potentially result in an unintended change in the valve setting due to exposure to an MRI procedure or other magnetic field. Based on reported complaints, this issue may have been a factor in up to 0.06% of units sold.
2. **The programming mechanism may not always operate properly.** If this occurs, it could potentially lead to an inability to modify the operating pressure of the valve with the hand-held programming tools, also referred to as the CODMAN[®] CERTAS[™] Therapy Management System (TMS). Based on our reported complaints, this issue may have been a factor in up to 0.4% of units sold.

Root Cause:

This recall is being initiated because our testing has shown that, in a small percentage of valves, the conditions noted above may exist for products released. In-depth root cause analysis is not completed at the present time.

Action Required:

- Acknowledge receipt of this Field Safety Notice to verify that you have been made aware of this possible issue and report unused items to be returned.
- Return any unused items to your local Codman offices as requested by this notice.
- Return the acknowledgement form to your Codman sales representative or local Codman office.

Advice for Physicians Treating Patients with Implanted CODMAN® CERTAS™ Programmable Valves:

1. **Mitigation Steps Regarding the MRI resistance feature:** Please follow the Instructions For Use (IFU) which recommend that the settings of implanted valves are verified following exposure to an MRI. We urge you to complete this step after every MRI.
2. **Mitigation Steps Regarding the programming mechanism difficulty:** In most cases, programming could be completed with additional attempts to re-program using the Therapy Management System(TMS) device and verification of setting, as described in the IFU. If additional attempts are unsuccessful, please contact your Codman Neuro representative who can provide further support.

Corrective Action

We are in the process of taking corrective actions to prevent re-occurrence. We regret the need to undertake this action and we thank you for your cooperation.

Distribution of this Field Safety Notice:

Please share this information with any staff that may use these products, especially those following patients who have been treated with the CODMAN® CERTAS™ Programmable Valves. .

Contact Information:

Consult with your local sales representative if you have questions concerning this notice.

Authorized European Representative Contact Reference Person:

Donal Hempenstall

tel: 44.1344.871186

fax: 44.1344.324687

This FSCA has/will be sent to the appropriate Regulatory Agency.

Please provide this notice to any Neurosurgeons or other clinicians at your facility who are treating patients with an implanted CODMAN® CERTAS™ Programmable Valve.