FDA Home<sup>3</sup> Medical Devices<sup>4</sup> Databases<sup>5</sup>

Class 2 Device Recall DeRoyal (R) Cardiac Hypothermia Tray

510(k)<sup>7</sup>[Registration & Listing<sup>8</sup>[Adverse Events<sup>9</sup>]Recalls<sup>10</sup>[PMA<sup>11</sup>]Classification<sup>12</sup>[Standards<sup>13</sup>]Inspections<sup>14</sup> CFR Title 21 15 Radiation-Emilting Products 18 X-Ray Assembler 17 Medsun Reports 18 CLIA 19 TPLC 20

New Search

Back to Search Results

Class 2 Recall DeRoyal (R) Cardiac Hypothermia Tray



Date Posted

July 29, 2014

Recall Status<sup>1</sup>

Open

Recall Number

Z-2125-2014

Recall Event ID

6874422

Product

DeRoyal (R) Cardiac Hypothermia Tray, REF 50-9422.07, 1 Per Pack, Rx Only STERILE EO, Distributed by: DeRoyal Industries, Inc. 200 DeBusk Lane, Powell,

TN 37849, USA; Made in Guatemala

Code Information

Lot Numbers: 30375680, 30515347, 30927691, 32922641, 33656541

Recalling Firm/

Manufacturer

DeRoyal Industries Inc

200 Debusk Ln

Powell, Tennessee 37849-4703

For Additional Information Contact

Tracy Edmundson 865-362-2334

Manufacturer Reason

for Recall

The firm distributed surgical kits which contained Irrigation Sets which were subsequently

recalled by Hospira.

Action

The recall was initiated by letter delivered via UPS on 6/4/2014. The firm requested that the consignee contact Stericycle to arrange for return and replacement of the product. A second notice was sent on 7/22/2014 to those accounts who did not respond to the initial notice. Distributors were requested to notify their end-users or provide a list of end-users to DeRoyal for direct notification. Effectiveness checks will be sent via email for those customers who have provided an email contact and via UPS for those customers who have not responded.

Quantity in Commerce

230 units

Distribution

Distributed in the states of MI, GA, OH, IL, MO, and IN

## Links on this page:

- 1. http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdamain
- http://www.addthis.com/bookmark.php
- 3. http://www.fda.gov/default.htm
- 4. http://www.fda.gov/MedicalDevices/default.htm
- 5. http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Databases/default.htm
- /scripts/cdrh/devicesatfda/index.cfm
- /scripts/cdrh/cfdocs/cfPMN/pmn.cfm
- 8. /scripts/cdrh/cfdocs/cfRL/rl.cfm
- 9. /scripts/cdrh/cfdocs/cfMAUDE/TextSearch.cfm

For details about termination of a recall see <u>Code of Federal Regulations (CFR) Title 21 §7.55</u><sup>23</sup>